

AD_____

REPORT NO T97-2

ACCURACY OF A COMMERCIALY AVAILABLE
TELEMETRY SYSTEM TO MEASURE CORE TEMPERATURE
DURING EXERCISE WHEN WEARING CHEMICAL
PROTECTIVE CLOTHING

**U S ARMY RESEARCH INSTITUTE
OF
ENVIRONMENTAL MEDICINE
Natick, Massachusetts**

JANUARY 1997



19971121 108

Approved for public release: distribution unlimited. DTIC QUALITY INSPECTED 3

**UNITED STATES ARMY
MEDICAL RESEARCH & DEVELOPMENT COMMAND**

DISCLAIMER

The views, opinions and/or findings in this report are those of the authors, and should not be construed as an official Department of the Army position, policy or decision, unless so designated by other official documentation.

Human subjects participated in these studies after giving their free and informed voluntary consent. Investigators adhered to Army Regulation 70-25 and USAMRMC Regulation 70-25 on Use of Volunteers in Research.

Citations of commercial organizations and trade names in this report do not constitute an official Department of the Army endorsement or approval of the products or services of these organizations.

DTIC AVAILABILITY NOTICE

Qualified requesters may obtain copies of this report from Commander, Defense Technical Information Center (DTIC) 8725 John J. Kingman Rd., STE 0944, Ft. Belvoir, VA. 22060-6218.

Approved for public release; distribution is unlimited.

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.				
1. AGENCY USE ONLY (Leave blank)		2. REPORT DATE January 1997		3. REPORT TYPE AND DATES COVERED Technical Report
4. TITLE AND SUBTITLE Accuracy of a Commercially Available Telemetry System to Measure Core Temperature During Exercise When Wearing Chemical Protective Clothing			5. FUNDING NUMBERS	
6. AUTHOR(S) M.A. Kolka, L. Levine and L.A. Stephenson				
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) U.S. Army Research Institute of Environmental Medicine Kansas Street Natick, MA 01760-5007			8. PERFORMING ORGANIZATION REPORT NUMBER T97-2	
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) U.S. Army Medical Research and Materiel Command Fort Detrick Frederick, MD 21702-5012			10. SPONSORING/MONITORING AGENCY REPORT NUMBER	
11. SUPPLEMENTARY NOTES				
12a. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release; distribution is unlimited.			12b. DISTRIBUTION CODE	
13. ABSTRACT (Maximum 200 words) Core temperature was measured by an esophageal thermistor and by an ingestible telemetry pill (Cortemp™, Human Technologies Inc., St. Petersburg, FL) during moderate exercise in women wearing a chemical protective clothing ensemble with a high thermal resistance (R_T , $0.4 \text{ m}^2 \cdot \text{K}^{-1} \cdot \text{W}^{-1}$). Telemetry pills were calibrated against a Hewlett Packard™ quartz thermometer the day prior to ingestion. Telemetry pills were ingested two hours before testing; pill ingestion was followed by a light meal. The esophageal thermistor was inserted to a length that was 25% of the standing height and adjusted to a previously determined "hot spot". Resting esophageal temperature (T_{es}) averaged $37.11 \pm 0.21^\circ\text{C}$ and resting pill temperature (T_{pill}) averaged $37.17 \pm 0.27^\circ\text{C}$. The combination of exercise ($225 \pm 30 \text{ W} \cdot \text{m}^{-2}$), clothing and ambient temperature ($T_a = 30^\circ\text{C}$) caused T_{es} to increase to an average of $38.67 \pm 0.28^\circ\text{C}$ and T_{pill} to increase to an average of $38.71 \pm 0.33^\circ\text{C}$ during the hour of treadmill walking. The two temperature measurements were compared by least squares regression techniques. For the eight individual experiments run, the regression coefficient (r) averaged 0.98 ± 0.01 . The calibrated temperature sensor ingested in these experiments provided accurate, usable core temperature data during logistically difficult experimental conditions.				
14. SUBJECT TERMS Telemetry; Thermoregulation; Protective Clothing; Heat Storage; Gender			15. NUMBER OF PAGES 25	
			16. PRICE CODE	
17. SECURITY CLASSIFICATION OF REPORT Unclassified	18. SECURITY CLASSIFICATION OF THIS PAGE Unclassified	19. SECURITY CLASSIFICATION OF ABSTRACT Unclassified	20. LIMITATION OF ABSTRACT	

TECHNICAL REPORT

NO. T97-2

**ACCURACY OF A COMMERCIALY AVAILABLE
TELEMETRY SYSTEM TO MEASURE CORE TEMPERATURE
DURING EXERCISE WHEN WEARING CHEMICAL
PROTECTIVE CLOTHING**

by

Margaret A. Kolka, Leslie Levine and Lou A. Stephenson

January 1997

U.S. Army Research Institute of Environmental Medicine

Natick, Massachusetts 01760-5007

CONTENTS

LIST OF FIGURES	iv
LIST OF TABLES	iv
FOREWORD	v
ACKNOWLEDGMENTS	vii
EXECUTIVE SUMMARY	1
INTRODUCTION	2
METHODS	4
RESULTS AND DISCUSSION	7
CONCLUSIONS	15
REFERENCES	16
DISTRIBUTION LIST	19

LIST OF FIGURES

- Figure 1: Esophageal temperature (T_{es}) and telemetered sensor temperature (T_{pill}) during experiments 1-4 in women wearing a relatively impermeable clothing ensemble during treadmill exercise.
- Figure 2: Esophageal temperature (T_{es}) and telemetered sensor temperature (T_{pill}) during experiments 5-8 in women wearing a relatively impermeable clothing ensemble during treadmill exercise.
- Figure 3: The linear relationship between esophageal temperature (T_{es} , X axis) and telemetered sensor temperature (T_{pill} , Y axis) during experiments 1-4 in women wearing a relatively impermeable clothing ensemble during treadmill exercise.
- Figure 4: The linear relationship between esophageal temperature (T_{es} , X axis) and telemetered sensor temperature (T_{pill} , Y axis) during experiments 5-8 in women wearing a relatively impermeable clothing ensemble during treadmill exercise.

LIST OF TABLES

- Table 1: Regression coefficients for the eight experiments comparing esophageal temperature (X variable) and pill temperature (Y variable).

FOREWORD

Heat exposure necessitates activation of thermoregulatory effector mechanisms to maintain core temperature. If core and skin temperatures increase, heat loss mechanisms are activated and increased sweat secretion and increased blood flow to the skin surface occur to eliminate body heat. The time an individual can tolerate work in the heat is decreased by the fabric barriers inherent in protective clothing systems. If heat transfer from the body surface via convection and evaporation is decreased, thermoregulatory heat balance is compromised. The rate of heat storage increases when subjects wear chemical protective clothing systems because the micro-environment at the skin surface (high water vapor content) impairs evaporative heat exchange, and the physical characteristics of the chemical protective clothing limit dry or sensible heat transfer.

There is always potential for heat stress in soldiers wearing chemical protective clothing. For this reason an unobtrusive and reliable measurement of core temperature has been sought by USAMRMC for monitoring human safety. USAMRMC has historically provided funding for telemetry research (Redmond *et al.*, 1992). The measurement of deep core temperature for research purposes or clinical safety is generally done using either esophageal or rectal temperature. Both temperatures are reproducible and not biased by environmental temperature. Telemetry systems offer a means to monitor core temperature using a swallowed "pill" as a sensor which transmits temperature information outside of the body. This technology is useful, especially for monitoring the safety of subjects in situations where, either clothing fully encapsulates the subject, prolonged

monitoring is required or hardwiring is not possible. This research was done to determine the accuracy of a temperature sensor and telemetry system as compared to esophageal temperature. The comparison was made from core temperature data collected from exercising women who were dressed in chemical protective clothing.

ACKNOWLEDGMENTS

This work would not have been possible without the volunteers who participated in the tests described in this report. We wish to thank them for their time and effort. We especially thank Dr. C. Gabarée, J. De Luca, B.S. Cadarette and B. Mair for their contributions to these studies.

EXECUTIVE SUMMARY

Core temperature was measured by an esophageal thermistor and by an ingestible telemetry pill (Cortemp™, Human Technologies Inc., St. Petersburg, FL) during moderate exercise in women wearing a chemical protective clothing ensemble with a high thermal resistance (R_T , $0.4 \text{ m}^2 \cdot \text{K}^{-1} \cdot \text{W}^{-1}$). Telemetry pills were calibrated against a Hewlett Packard™ quartz thermometer the day prior to ingestion. Telemetry pills were ingested two hours before testing; pill ingestion was followed by a light meal. The esophageal thermistor was inserted to a length that was 25% of the standing height and adjusted to a previously determined “hot spot”. Resting esophageal temperature (T_{es}) averaged $37.11 \pm 0.21^\circ\text{C}$ and resting pill temperature (T_{pill}) averaged $37.17 \pm 0.27^\circ\text{C}$. The combination of exercise ($225 \pm 30 \text{ W} \cdot \text{m}^{-2}$), clothing and ambient temperature ($T_a = 30^\circ\text{C}$) caused T_{es} to increase to an average of $38.67 \pm 0.28^\circ\text{C}$ and T_{pill} to increase to an average of $38.71 \pm 0.33^\circ\text{C}$ during the hour of treadmill walking. The two temperature measurements were compared by least squares regression techniques. For the eight individual experiments run, the regression coefficient (r) averaged 0.98 ± 0.01 . The calibrated temperature sensor ingested in these experiments provided accurate, usable core temperature data during logistically difficult experimental conditions.

INTRODUCTION

The measurement of core temperature for research purposes or clinical safety is generally done using either esophageal or rectal temperature. Both temperatures are reproducible and not biased by environmental temperature (Breglemann, 1987; Gerbrandy *et al.*, 1954; Mead *et al.*, 1949); however, the slow response time of rectal temperature is well documented (Eichna *et al.*, 1951; Gerbrandy *et al.*, 1954; Kolka *et al.*, 1987; Mittleman *et al.*, 1988; Molnar *et al.*, 1974). Esophageal temperature is an ideal method for tracking blood temperature in the right heart as it most accurately reflects the integrated temperature of the pre-optic/anterior hypothalamus. Consequently, it responds very quickly to dynamic changes in mean body temperature (Breglemann, 1987; Gerbrandy *et al.*, 1954; Rowell, 1983; Shiraki *et al.*, 1986; Shiraki *et al.*, 1988). Telemetry systems offer a means to monitor core temperature using a swallowed "pill" as a sensor which transmits temperature information to a receiver outside of the body.

Chemical protective overgarments, fire protective garments, and toxic site cleanup garments worn by military or civilian personnel create a logistically difficult situation in which to safely and accurately monitor the core temperature of these individuals. In these workers, heat strain is a major problem as core temperature continues to increase because the heat produced by muscular work and/or environmental conditions cannot be dissipated through the clothing system. Telemetry technology is useful, especially for monitoring the safety of subjects in situations where, either clothing fully encapsulates the subject, prolonged monitoring is required or hardwiring is not possible (Mackay, 1970; Wolff,

1961; Redmond *et al.*, 1992). Telemetry technology has been used previously to measure core temperature in human subjects in various experimental protocols (Fox *et al.*, 1961; Gibson *et al.*, 1981; Mittal *et al.*, 1991; Sparling *et al.*, 1993; Stephenson *et al.*, 1992; Kolka *et al.*, 1993).

STATEMENT OF PURPOSE

The purpose of this study was to compare how well a commercially available temperature telemetry system tracked rapid changes in core temperature compared with esophageal temperature during exercise in subjects wearing a chemical protective clothing ensemble which limited both evaporative heat transfer and dry heat transfer, allowing continued heat storage during the experiment.

METHODS

Four women volunteered to serve as test subjects after they were informed of the purpose, procedures, and known risks of this study. Each signed a consent form describing the study which was approved by appropriate human use review committees. The average (\pm SD) age was 25.6 ± 9.6 yr, height was 1.68 ± 0.05 m, mass was 63.0 ± 11.0 kg, DuBois surface area was 1.71 ± 0.15 m² and maximal aerobic power was 2.66 ± 0.30 L \cdot min⁻¹.

Two methods were used to measure core temperature: esophageal temperature (T_{es}), and pill temperature (CorTemp™, Human Technologies, Inc., St. Petersburg, FL ingestible temperature sensor). Esophageal temperature was measured at a pre-determined location in the esophagus deemed the "hot spot" using a calibrated thermistor. The "hot spot" was located by systematically determining temperature in the esophagus beginning at approximately 25% of each volunteer's height until the temperature was at a peak, presumably due to the proximity to the great vessels of the heart. The CorTemp™ telemetry system included an ingestible temperature sensor (2 cm by 1.3 cm in diameter), FM antenna and data recorder system. The calibrated temperature sensor included a silver oxide battery (1.5 V) which provided the power for sensing and transmitting temperature. The components of the sensors were encapsulated in epoxy and covered with silicone rubber. Temperature was transmitted through the body to a double bandoleer-type antenna and recorded by a data logger.

All exercise testing occurred between 0730 and 0900 h. Each subject was tested on 1-3 occasions. Upon arrival at the laboratory the volunteer swallowed the telemetry pill with water and ate a light breakfast. Two (± 0.5) hours after swallowing the pill, the volunteer was taken to the environmental test chamber and inserted the esophageal temperature probe.

After instrumentation, each subject dressed in chemical protective clothing (modified MOPP 4: BDU, overgarment, overboots, hood, gloves, open mask for metabolic rate measurements and to accommodate the esophageal thermocouple). After complete instrumentation and dressing, the antenna for the telemetry system was attached to the subject's torso, and then she sat in a chair which was positioned on the treadmill. After equilibration with the environment (15-30 min), treadmill exercise began and continued for 60 minutes. The ambient temperature in the environmental test chamber was 30°C and the dew point temperature was 11.5°C. Esophageal temperature and pill temperature were measured every 30 s throughout the experiment. Heart rate was measured at 5 min intervals by electrocardiography.

Subject Safety

All of the procedures in this study fell within the framework, restrictions and safety limitations of the USARIEM Type Protocol for Human Research Studies in the areas of

Thermal, Hypoxic and Operational Stress, Exercise, Nutrition and Military Performance, which was in effect at the time when the research was done.¹

STATISTICAL ANALYSES

Esophageal temperature and pill temperature were compared at rest and during exercise by analysis of variance procedures with repeated measures. In addition, esophageal temperature and pill temperature data were compared by least squares regression techniques.

¹Approved 14 Dec 1994. The type protocol provides information and explanations about conditions, standards and safeguards, in order to serve as an encompassing framework for specific in-house studies in its general subject area. It is to be used as a reference to facilitate the understanding and review of specific study protocols which conform to its provisions, and thus do not exceed the degree of risk, and safety limits herein stipulated (reference para 18, USAMRMC Reg 70-25).

RESULTS AND DISCUSSION

A published report (Kolka *et al.*, 1993) documented the limitations of the telemetry technique during laboratory experiments. The pitfalls in the use of two different telemetry methods for monitoring core temperature (Stephenson *et al.*, 1992) included capturing telemetry data (experimental telemetry), slow responding temperature sensor (experimental telemetry) and movement of the sensor (experimental and CorTemp™ telemetry) in the gastrointestinal system. The promising results using the CorTemp™ system in the previous study (Stephenson *et al.*, 1992) prompted us to continue testing this device in a more difficult scenario. The data from the current study suggest that under controlled experimental conditions, this commercially available, off-the-shelf system, can be used for effective, accurate core temperature measurements in encapsulated human volunteers.

The primary purpose of this study was to evaluate the data from a temperature pill telemetry system and determine whether the data provided an accurate index of core temperature in fully encapsulated subjects. Experiments were run under severe conditions created by a combination of exercise intensity, clothing with a high thermal resistance and low evaporative capacity, and warm environmental conditions. This combination caused core temperature to increase $\sim 1.5^{\circ}\text{C}$ in 60 minutes, during which time skin temperature remained above 36°C and heart rate approached $170 \text{ b}\cdot\text{min}^{-1}$. Resting esophageal temperature averaged $37.11 \pm 0.21^{\circ}\text{C}$ and resting pill temperature averaged $37.17 \pm 0.27^{\circ}\text{C}$. The combination of exercise ($225 \pm 30 \text{ W}\cdot\text{m}^{-2}$), clothing and ambient temperature ($T_a =$

30°C) increased esophageal temperature to an average of $38.67 \pm 0.28^\circ\text{C}$ and pill temperature to an average of $38.71 \pm 0.33^\circ\text{C}$ during the hour of treadmill walking. The pattern of esophageal temperature and pill temperature for all eight experiments is shown in Figures 1 and 2. In many of the experiments (Figures 1 and 2) there are aberrant data points from the telemetry sensors. These occurred because of difficulties with the antennae and receivers, not because telemetry sensor signal was lost. Improvements in reception of telemetered signal have been made and are currently ongoing. The comparison of the two temperature measurements by least squares regression techniques (r) is shown in Figures 3 and 4. For the eight individual experiments run, the r averaged 0.98 ± 0.01 (Table 1).

Esophageal temperature was measured in this study because it responds rapidly to changes in body temperature (Breglemann, 1987; Gerbrandy *et al.*, 1954; Rowell, 1983; Shiraki *et al.*, 1986; Shiraki *et al.*, 1988). Tympanic or auditory meatus temperature has been used as an index of internal body temperature in laboratory studies (Baker *et al.*, 1972; Greenleaf *et al.*, 1972; Sharkey *et al.*, 1987; Shiraki *et al.*, 1986; Shiraki *et al.*, 1988). The main detractor for using tympanic temperature and the reason this index was not used in the current study, is that tympanic temperature more reliably tracks skin temperature than core temperature in humans (Greenleaf *et al.*, 1972; Sharkey *et al.*, 1987; Shiraki *et al.*, 1986). Initially, tympanic temperature was used because it responded similarly to brain temperature in the monkey and the cat (Baker *et al.*, 1972), and it was proposed that tympanic temperature was the best core temperature index in humans

because it more accurately reflects hypothalamic temperature and would also account for selective brain cooling (Brinnet *et al.*, 1989; Cabanac, 1986; Cabanac *et al.*, 1979).

However, the existence of selective brain cooling in humans has not been proven and remains highly controversial (Cabanac, 1986; Nadel, 1987; Wenger, 1987; Jessen and Kuhnen, 1992).

The environmental conditions imposed on the subjects in the study limited dry heat flux between the skin and the clothing, and limited dry heat flux through the clothing. Heat loss in warm environments during exercise occurs by evaporative heat loss, but the relatively non-porous clothing worn by our subjects resisted transmission of water vapor. In these experiments, all subjects continued to store heat during exercise with no leveling off of internal temperature (Figures 1 and 2). It is clear that T_{pill} (Table 1) is an accurate index of core temperature under the conditions of this study.





Figure 1: Esophageal temperature (T_{es}) and telemetered sensor temperature (T_{pill}) during experiments 1-4 in women wearing a relatively impermeable clothing ensemble during treadmill exercise.





Figure 2: Esophageal temperature (T_{es}) and telemetered sensor temperature (T_{piil}) during experiments 5-8 in women wearing a relatively impermeable clothing ensemble during treadmill exercise.







Figure 3: The linear relationship between esophageal temperature (T_{es} , X axis) and telemetered sensor temperature (T_{pill} , Y axis) during experiments 1-4 in women wearing a relatively impermeable clothing ensemble during treadmill exercise.





Figure 4: The linear relationship between esophageal temperature (T_{es} , X axis) and telemetered sensor temperature (T_{pill} , Y axis) during experiments 5-8 in women wearing a relatively impermeable clothing ensemble during treadmill exercise.

Table 1. Regression coefficients for the eight experiments comparing esophageal temperature (X variable) and pill temperature (Y variable).

<u>Experiment</u>	<u>r</u>	<u>slope</u>	<u>Y-intercept</u>
1	0.99	0.980	0.674
2	0.98	0.828	6.401
3	0.97	0.816	7.020
4	0.97	0.879	4.512
5	0.98	0.885	4.612
6	0.99	0.877	4.850
7	0.98	1.103	-3.948
8	0.99	1.140	-5.147

CONCLUSIONS

In this study, pill temperature was not different than esophageal temperature during exercise, and the total change in pill temperature during the 60 min experiment was not significantly different than the change in esophageal temperature. In conclusion, the telemetry system as a core temperature data acquisition system was reliable, provided that preliminary screening by water bath calibration eliminated those sensors (pills) which measured temperature inaccurately. For the most part, the telemetry system accurately measured core temperature. The concept of using a temperature sensor in a pill is useful (Wolff, 1961) if care is taken to 1) calibrate and test the sensor before ingestion; 2) locate the sensor, through time of ingestion, in the appropriate portion of the gastrointestinal tract useful for the specific protocol; a small meal given with ingestion can aid passing the pill out of the stomach; and 3) monitor receiver location so that all data are collected.

REFERENCES

- Baker M.A., Stocking R.A., and Meehan J.P. (1972) Thermal relationship between tympanic membrane and hypothalamus in conscious cat and monkey. Journal of Applied Physiology 32, 739-742.
- Brengelmann G.L. (1987) Dilemma of body temperature measurement. In: Man in Stressful Environments. Thermal and Work Physiology. edited by K. Shiraki, and M. K. Yousef. Springfield, IL: Charles C. Thomas, pp. 5-22.
- Brinnel H., and Cabanac M. (1989) Tympanic temperature is a core temperature in humans. Journal of Thermal Biology 14, 47-53.
- Cabanac M. (1986) Keeping a cool head. News in Physiological Sciences 1, 41-44.
- Cabanac M., and Caputa M. (1979) Natural selective cooling of the human brain: evidence of its occurrence and magnitude. Journal of Physiology (London) 286, 255-264.
- Eichna L.W., Berger A.R., Rader B., and Becker W.H. (1951) Comparison of intracardiac and intravascular temperatures with rectal temperatures in man. Journal of Clinical Investigation 30, 353-359.
- Fox R.H., Goldsmith R. and Wolff H.S. (1961) The use of a radio pill to measure deep body temperature. Journal of Physiology (London) 160, 22-23.
- Gerbrandy J., Snell E.S., and Cranston W.I. (1954) Oral, rectal and oesophageal temperatures in relation to central temperature control in man. Clinical Sciences 13, 615-624.
- Gibson T.M., Redman P.J. and Belyavin A.J. (1981) Prediction of oesophageal temperatures from core temperatures measured at other sites in man. Clinical Physical and Physiological Measures 2, 247-256.
- Greenleaf J.E., and Castle B.L. (1972) External auditory canal temperature as an estimate of core temperature. Journal of Applied Physiology 32, 194-198.
- Jessen C., and Kuhn G. (1992) No evidence for brain stem cooling during face fanning in humans. Journal of Applied Physiology 72, 664-669.
- Kolka M.A., Stephenson L.A., Bruttig S.P., Cadarette B.S., and Gonzalez R.R. (1987) Human thermoregulation after atropine and/or pralidoxime administration. Aviation, Space and Environmental Medicine 58, 545-549.

- Kolka M.A., Quigley M.D., Blanchard L.A., Toyota D.A., and Stephenson L.A. (1993) Validation of a temperature telemetry system during moderate and strenuous exercise. Journal of Thermal Biology 18:203-210.
- Mackay R.S. (1970) Bio-Medical Telemetry. Second Ed., New York: John Wiley & Sons, Inc., pp. 154-159.
- Mead J., and Bonmarito C.L. (1949) Reliability of rectal temperatures as an index of internal body temperature. Journal of Applied Physiology 2, 97-109.
- Mittal B.B., Sathiaselan V., Rademaker A.W., Pierce M.C., Johnson P.M. and Brand W.N. (1991) Evaluation of an ingestible telemetric temperature sensor for deep hyperthermia applications. International Journal of Radiation Oncology in Biology and Physiology 21, 1353-1361.
- Mittleman K.D., and Mekjavic I.B. (1988) Effect of occluded venous return on core temperature during cold water immersion. Journal of Applied Physiology 65, 2709-2713.
- Molnar G.W., and Read R.C. (1974) Studies during open-heart surgery on the special characteristics of rectal temperature. Journal of Applied Physiology 36, 333-336.
- Nadel E.R. (1987) Comments on "Keeping a cool head" by M. Cabanac. News in Physiological Sciences 2, 33.
- Redmond D., Leu, J.C., Popp K.A., Gutierrez P.E. and Hall S.W., Jr. (1990) The Walter Reed temperature pill: Evaluations for clinical and investigational usage. Walter Reed Army Institute of Research Technical Report.
- Rowell L.B. (1983) Cardiovascular adjustments to thermal stress. In: Handbook of Physiology-The Cardiovascular System III, edited by J. T. Shepherd and F. M. Abboud. Bethesda: Am. Physiol. Soc., Chapter 27, pp.967-1023.
- Sharkey A., Elliott P., Lipton J.M., and Giesecke A.H. (1987) The temperature of the air within the external auditory meatus compared with esophageal temperature during anaesthesia and surgery. Journal of Thermal Biology 12, 11-13.
- Shiraki K., Konda N., and Sagawa S. (1986) Esophageal and tympanic temperature responses to core blood temperature changes during hyperthermia. Journal of Applied Physiology 61, 98-102.
- Shiraki K., Sagawa S., Tajima F., Yokota A., Hashimoto M., and Brengelmann G.L. (1988) Independence of brain and tympanic temperatures in an unanesthetized human. Journal of Applied Physiology 65, 482-486.

Sparling P.B., Snow T.K., and Millard-Stafford M. (1993) Monitoring core temperature during exercise: ingestible sensor vs rectal thermistor. Aviation, Space and Environmental Medicine 64, 760-763.

Stephenson L.A., Quigley M.D., Blanchard L. A., Toyota D.A. and Kolka M.A. (1992) Validation of two telemetry pill systems in humans during moderate and strenuous exercise. US Army Research Institute of Environmental Medicine Technical Report, T10\92.

Wenger C.B. (1987) More comments on "Keeping a cool head" by M. Cabanac. News in Physiological Sciences 2, 150.

Wolff H.S. (1961) The radio pill. New Scientist 16, 419-421.

DISTRIBUTION LIST

2 Copies to:

Defense Technical Information Center
8725 John J. Kingman Road, STE 0944
Fort Belvoir, VA 22060-6218

Office of the Assistant Secretary of Defense (Health Affairs)
ATTN: Medical Readiness, Pentagon
Washington DC 20301-0103

Commander
US Army Medical Research and Materiel Command
ATTN: MCMR-OP
Fort Detrick MD 21702-5012

Commander
U.S. Army Medical Research and Materiel Command
ATTN: MCMR-PLC
Fort Detrick MD 21702-5012

Commander
U.S. Army Medical Research and Materiel Command
ATTN: MCMR-PLE
Fort Detrick MD 21702-5012

Commandant
Army Medical Department Center and School
ATTN: HSMC-FR, Bldg. 2840
Fort Sam Houston TX 78236

1 Copy to:

Deputy Director for Medical Readiness, Joint Chiefs of Staff, J4
ATTN: J4-MRD
4000 Joint Staff, Pentagon
Washington DC 20318-4000

HQDA
Assistant Secretary of the Army
(Research, Development and Acquisition)
ATTN: SARD-TM, Pentagon
Washington DC 20316-0103

HQDA
Office of the Surgeon General
Assistant Surgeon General
ATTN: DASG-RDZ/Executive Assistant Surgeon General
Room 3E368, Pentagon
Washington DC 20310-2300

HQDA
Office of the Surgeon General
ATTN: DASG-ZA
5109 Leesburg Pike
Falls Church VA 22041-3258

HQDA
Office of the Surgeon General
ATTN: DASG-DB
5109 Leesburg Pike
Falls Church VA 22041-3258

HQDA
Office of the Surgeon General
ATTN: DASG-MS
5109 Leesburg Pike
Falls Church VA 22041-3258

HQDA
Office of the Surgeon General
Preventive Medicine Consultant
ATTN: SGPS-PSP
5109 Leesburg Pike
Falls Church VA 22041-3258

Uniformed Services University of the Health Sciences
4301 Jones Bridge Road
Bethesda MD 20814-4799

Uniformed Services University of the Health Sciences
ATTN: Department of Preventative Medicine
4301 Jones Bridge Road
Bethesda MD 20814-4799

Commandant
Army Medical Department Center & School
ATTN: Chief Librarian Stimson Library
Bldg 2840, Room 106
Fort Sam Houston TX 78234-6100

Commandant
Army Medical Department Center & School
ATTN: Director of Combat Development
Fort Sam Houston TX 78234-6100

Commander
U.S. Army Aeromedical Research Laboratory
ATTN: MCMR-UAX-SI
Fort Rucker AL 36362-5292

Commander
U.S. Army Medical Research Institute of Chemical Defense
ATTN: MCMR-UVZ
Aberdeen Proving Ground MD 21010-5425

Commander
U.S. Army Medical Materiel Development Activity
ATTN: MCMR-UMZ
Fort Detrick MD 21702-5009

Commander
U.S. Army Institute of Surgical Research
ATTN: MCMR-USZ
Fort Sam Houston TX 78234-5012

Commander
U.S. Army Medical Research Institute of Infectious Diseases
ATTN: MCMR-UIZ-A
Fort Detrick MD 21702-5011

Director
Walter Reed Army Institute of Research
ATTN: MCMR-UWZ-C (Director for Research Management)
Washington DC 20307-5100

Commander
U.S. Army Soldier's Systems Command
ATTN: ANSSC-CG
Natick MA 01760-5000

Commander
U.S. Army Natick Research, Development & Engineering Center
ATTN: SSCNC-Z
Natick MA 01760-5000

Commander
U.S. Army Natick Research, Development & Engineering Center
ATTN: SSCNC-T
Natick MA 01760-5002

Commander
U.S. Army Natick Research, Development & Engineering Center
ATTN: SSCNC-S-IMI
Natick MA 01760-5040

Commander
U.S. Army Natick Research, Development & Engineering Center
ATTN: SSCNC-TM
U.S. Marine Corps Representative
Natick MA 01769-5004

Commander
U.S. Army Research Institute for Behavioral Sciences
5001 Eisenhower Avenue
Alexandria VA 22333-5600

Commander
U.S. Army Training and Doctrine Command
Office of the Surgeon
ATTN: ATMD
Fort Monroe VA 23651-5000

Commander
U.S. Army Center for Health Promotion and Preventative Medicine
Aberdeen Proving Ground MD 21010-5422

Director, Biological Sciences Division
Office of Naval Research - Code 141
800 N. Quincy Street
Arlington VA 22217

Commanding Officer
Naval Medical Research & Development Command
NNMC/Bldg 1
Bethesda MD 20889-5044

Commanding Officer
U.S. Navy Clothing & Textile Research Facility
ATTN: NCTRF-01, Bldg 86
Natick MA 01760-5000

Commanding Officer
Navy Environmental Health Center
2510 Walmer Avenue
Norfolk VA 23513-2617

Commanding Officer
Naval Aerospace Medical Institute (Code 32)
Naval Air Station
Pensacola FL 32508-5600

Commanding Officer
Naval Medical Research Institute
Bethesda MD 20889

Commanding Officer
Naval Health Research Center
P.O. Box 85122
San Diego CA 92138-9174

Commander
Armstrong Medical Research Laboratory
Wright-Patterson Air Force Base OH 45433

U.S. Airforce Aeromedical Library
Document Services Section
2511 Kennedy Circle
Brooks Air Force Base TX 78235-5122

Commander
USAF School of Aerospace Medicine
Brooks Air Force Base TX 78235-5000

Director
U.S. Army Research Laboratory
Human Research & Engineering Directorate
Aberdeen Proving Ground MD 21005-5001

Commander
U.S. Army Military History Institute
ATTN: Chief, Historical Reference Branch
Carlisle Barracks
Carlisle PA 17013-5008

U.S. Army Biomedical R&D Representative for Science
and Technology Center, Far East
ATTN: AMC-S&T, FE Unit 45015
APO AP 96343-5015

Director
Defence and Civil Institute of Environmental Medicine
1133 Sheppard Avenue W.
P.O. Box 200
Downsview, Ontario Canada M3M 3B9

Department of National Defence, CANADA
Defence Research Establishment Ottawa
ATTN: Head Physiology Group
Environmental Protection Section, Protective Sciences Division
Ottawa, Ontario Canada K1A 0Z4

Head, Environmental Physiology Section
Defence and Civil Institute of Environmental Medicine
1133 Sheppard Avenue W.
P.O. Box 200
Downsview, Ontario Canada M3M 3B9

National Defence Headquarters, CANADA
ATTN: Research and Development Branch
Human Performance Directorate
305 Rideau Street
Ottawa Ontario Canada K1AOK2